

JAN 13 2011  
2011  
3/9/11

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

mailed  
1/10/11  
VE

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130179 ✓

Name V. Eugene Fowler ID# 4386 ✓ Date 1/8/11 ✓

A Agency Alaska State Troopers Phone # 262-4453

Instrument Location Soldotna AST Post

B Alco S/N X12446 ✓ Target Value .084 ✓ High Pressure 1000

Alco Test Value Average .084 ✓ .084 ✓  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature V. E. Fowler

(OVER)

bmb  
2/8/11

I, Nita J. Bolz, after being first duly sworn, depose and state under the area below)


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 11<sup>th</sup> day of MAR, 2011 AN

(Notary Seal Stamp)

Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130179

(CONTINUED FROM FRONT PAGE)

- C. TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130179

JANUARY 08, 2011

OPERATOR'S NAME:

FOWLER/4/E

OPERATOR'S NUMBER: 4386

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: SLT5

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .084

ALCO S/N: X124646

### — BREATH ANALYSIS —

.084 ADJUSTED FOR 30.85 in		
ALCO TARGET	.086	19:20
BLANK TEST	.000	19:20
INTERNAL STANDARD	VERIFIED	19:21
ALCO TO 30.85 in	.084	19:21
BLANK TEST	.000	19:22
SUBJECT SAMPLE	.000	19:23
BLANK TEST	.000	19:23
ALCO TO 30.85 in	.084	19:24
BLANK TEST	.000	19:25

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130179

JANUARY 08, 2011

TIME 19:13

### — DIAGNOSTIC CHECK —

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	42c
BAROMETER:	30.85 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNPOQRSTUVWXYZ[\]^\_`abcdefgijklmno  
pqrstuvwxyz{|}~■